

PG7601 w/ AMH

When GAUGE/ABSOLUTE mode is selected, Bell jar, ref port, Barom, volume and neg side Of DUT are connected. When VACUUM MODE Is selected, everything is connected and line P and offset are set. There is an isolation valve on the volume in case you do not need to use it, or to help set the pressure when determining the offset.

Measured Pressure +

Measured Pressure – (ref side)

Supply Pressure - 10% above Test Pressure

AMH Drive Pressure – 40 psi

Control Vacuum Source



