



CERTIFICATE OF ANALYSIS

FLUKE ELECTRONICS CORPORATION  
4765 EAST BEAUTIFUL LANE

PHOENIX, AZ 85044

Product.....: EL DEAD WEIGHT TESTER FLUID-22

Customer Purchase Order: 569929  
Order Number.....: 95359  
Shipment Date.....: 07-20-15  
Batch Number.....: 10314887U007  
Lab Sample Number.....: 1112-Q0258

	Acceptable Limits		Test
	minimum	maximum	results
VISCOSITY INDEX (ASTM D445)			
cSt Vis. @ 40C...:	19.8	24.2	21.92
VISUAL INSPECTIONS			
Appearance.....:	CLEAR AMBER		OK
FT-IR FINGERPRINT SCAN (Oils)			
FT-IR Scan No.....:			OK